
Midmark M11 - v1.0.0
Total Cycles: 4031

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2023 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.7 F 0.0
2:00 143.9 F 2.2
4:00 161.7 F 5.5
6:00 185.0 F 9.1
8:00 211.5 F 13.3
10:00 217.3 F 6.8
12:00 230.6 F 11.2
14:00 231.9 F 7.2
16:00 241.5 F 12.6
18:00 240.2 F 10.8
20:00 252.4 F 16.8
22:00 263.1 F 23.0
24:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.0 F 29.3
2:00 272.3 F 29.1
3:00 272.4 F 29.6
4:00 272.1 F 29.1

Min 270.9 F 28.3
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 215.8 F 0.0
10:00 216.7 F 0.0
15:00 220.6 F 0.0
20:00 222.4 F 0.0
25:00 222.0 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:03:03
CYCLE COMPLETE

11/16/2023 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____