
Midmark M11 - v1.0.0
Total Cycles: 5455

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2025 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.6 F 0.0
2:00 177.3 F 3.2
4:00 191.0 F 7.5
6:00 203.4 F 13.4
8:00 218.5 F 7.2
10:00 232.3 F 12.8
12:00 236.1 F 10.0
14:00 236.5 F 8.0
16:00 249.4 F 15.2
18:00 261.8 F 22.3
20:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.1 F 28.9
2:00 271.8 F 29.1
3:00 271.8 F 28.7
4:00 272.2 F 29.3

Min 271.1 F 28.4
Max 272.3 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 217.4 F 0.0
10:00 231.0 F 0.0
15:00 227.0 F 0.0
20:00 227.0 F 0.0
25:00 229.6 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

03/27/2025 01:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____