
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002364
10:23 05/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:34
STERILIZING 4:45
136.4 C 224kPa 4:45
136.4 C 225kPa 5:15
136.4 C 224kPa 5:45
136.3 C 224kPa 6:15
136.1 C 222kPa 6:45
136.4 C 223kPa 7:15
136.4 C 225kPa 7:45
135.9 C 218kPa 8:15
135.9 C 218kPa 8:15
VENTING 8:15
AIR DRYING 9:12
CYCLE COMPLETE 9:19

Digital Signature #
FAD8CDF9BF707BF1
