
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006986
15:16 26/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:13
STERILIZING 4:17
136.4 C 220kPa 4:17
136.1 C 222kPa 4:47
136.4 C 220kPa 5:17
136.2 C 224kPa 5:47
136.2 C 220kPa 6:17
136.4 C 225kPa 6:47
136.3 C 224kPa 7:17
136.4 C 225kPa 7:47
136.4 C 225kPa 7:47
VENTING 7:47
AIR DRYING 8:43
CYCLE COMPLETE 8:46

Digital Signature #
E490A256998E886E
