

-----  
Midmark M11 - v1.0.5  
Total Cycles: 523

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2021 02:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.8 F 0.0  
2:00 117.4 F 1.2  
4:00 127.3 F 3.8  
6:00 154.4 F 7.9  
8:00 182.2 F 13.0  
10:00 211.2 F 7.7  
12:00 220.4 F 12.8  
14:00 233.9 F 9.9  
16:00 235.5 F 8.8  
18:00 246.1 F 14.0  
20:00 257.8 F 20.3  
22:00 267.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.8 F 30.7  
1:00 271.8 F 30.7  
2:00 271.8 F 30.8  
3:00 272.1 F 30.8  
4:00 272.6 F 30.8

Min 271.1 F 30.6  
Max 272.8 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 234.9 F 0.0  
10:00 228.5 F 0.0  
15:00 231.1 F 0.0  
20:00 225.6 F 0.0  
25:00 229.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:03  
HEATING: 23:56  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 25:53  
TOTAL CYCLE: 00:57:54

11/08/2021 02:57 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01