
Midmark M11 - v1.0.0
Total Cycles: 2138

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 182.8 F 3.1
4:00 189.0 F 7.5
6:00 208.3 F 12.4
8:00 220.3 F 6.5
10:00 232.8 F 11.6
12:00 232.5 F 8.1
14:00 242.6 F 11.8
16:00 244.3 F 12.4
18:00 256.5 F 18.8
20:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.8
2:00 272.0 F 28.8
3:00 272.1 F 28.6
4:00 272.1 F 29.0

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 231.6 F 0.0
10:00 231.7 F 0.0
15:00 227.2 F 0.0
20:00 227.3 F 0.0
25:00 229.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

07/15/2022 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____