
Midmark M11 - v1.0.0
Total Cycles: 2418

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.2 F 0.0
2:00 211.1 F 3.8
4:00 215.9 F 8.7
6:00 222.7 F 9.0
8:00 232.1 F 9.2
10:00 239.2 F 9.6
12:00 243.3 F 12.5
14:00 241.8 F 11.2
16:00 255.2 F 18.0
18:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.0 F 28.5
2:00 272.5 F 29.0
3:00 271.5 F 28.3
4:00 272.1 F 28.7

Min 271.4 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 232.9 F 0.0
10:00 230.2 F 0.0
15:00 227.3 F 0.0
20:00 225.2 F 0.0
25:00 225.3 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

10/12/2022 11:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____