
Midmark M11 - v1.0.5
Total Cycles: 3982

Name: RITTER2
Sterilizer ID: V2313228
10/01/2024 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 146.8 F 1.6
4:00 150.5 F 4.1
6:00 170.1 F 7.8
8:00 192.0 F 12.1
10:00 212.5 F 6.1
12:00 225.8 F 11.0
14:00 227.3 F 7.1
16:00 240.2 F 12.7
18:00 240.0 F 11.1
20:00 252.0 F 17.1
22:00 262.5 F 23.4
24:00 272.8 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.2
1:00 272.2 F 30.1
2:00 272.4 F 30.1
3:00 272.7 F 30.3
4:00 272.3 F 30.5

Min 271.8 F 30.0
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 210.1 F 0.0
10:00 230.7 F 0.0
15:00 228.3 F 0.0
20:00 212.0 F 0.0
25:00 219.0 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:08
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

10/01/2024 11:37 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____