
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006256
7:54 07/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:25
STERILIZING 4:26
136.4 C 222kPa 4:26
136.2 C 222kPa 4:56
136.1 C 222kPa 5:26
136.3 C 222kPa 5:56
136.1 C 218kPa 6:26
136.3 C 222kPa 6:56
136.1 C 219kPa 7:26
136.2 C 221kPa 7:56
136.2 C 221kPa 7:56
VENTING 7:56
AIR DRYING 8:45
CYCLE COMPLETE 8:53

Digital Signature #
F6E697C71CB7B85B
