
Midmark M11 - v1.0.0
Total Cycles: 4692

Name: RITTER 1
Sterilizer ID: V2178198
07/02/2024 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.7 F 0.0
2:00 139.3 F 1.6
4:00 148.1 F 4.4
6:00 174.5 F 7.5
8:00 198.0 F 11.1
10:00 215.8 F 4.2
12:00 218.5 F 7.7
14:00 229.6 F 11.8
16:00 230.0 F 7.5
18:00 241.9 F 12.7
20:00 239.7 F 10.5
22:00 251.5 F 16.2
24:00 261.9 F 22.1
26:00 271.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 271.9 F 29.0
2:00 272.2 F 28.9
3:00 271.1 F 28.4
4:00 272.4 F 29.2

Min 271.1 F 28.4
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 202.2 F 0.0
10:00 206.0 F 0.0
15:00 210.2 F 0.0
20:00 209.3 F 0.0
25:00 211.2 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

07/02/2024 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____