
Midmark M11 - v1.0.5
Total Cycles: 1151

Name: RITTER2
Sterilizer ID: V2313228
06/01/2022 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.2 F 0.0
2:00 186.5 F 3.2
4:00 195.7 F 8.0
6:00 214.5 F 12.7
8:00 226.5 F 9.1
10:00 235.2 F 9.9
12:00 242.5 F 13.0
14:00 242.6 F 12.1
16:00 255.5 F 18.8
18:00 266.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.5
2:00 271.6 F 29.2
3:00 272.4 F 29.6
4:00 271.5 F 29.0

Min 271.5 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 228.0 F 0.0
10:00 237.9 F 0.0
15:00 229.1 F 0.0
20:00 228.6 F 0.0
25:00 225.2 F 0.0
30:00 221.5 F 0.0

Drying Complete

FILLING: 1:55
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

06/01/2022 11:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____