
Midmark M11 - v1.0.5
Total Cycles: 2926

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 08:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 82.9 F 0.0
4:00 127.7 F 0.0
6:00 112.2 F 0.0
8:00 158.7 F 0.0
10:00 181.3 F 1.0
12:00 208.1 F 4.5
14:00 222.2 F 7.9
16:00 231.6 F 11.5
18:00 225.6 F 5.8
20:00 237.9 F 10.6
22:00 235.3 F 8.3
24:00 245.6 F 13.3
26:00 243.9 F 12.1
28:00 255.8 F 18.3
30:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.8
2:00 271.7 F 28.7
3:00 272.4 F 28.7
4:00 271.7 F 28.7

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 229.9 F 0.0
10:00 235.3 F 0.0
15:00 228.0 F 0.0
20:00 223.5 F 0.0
25:00 226.2 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

09/26/2023 09:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____