

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3386

Name: RITTER2  
Sterilizer ID: V2313228  
02/15/2024 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 76.6 F 0.0  
4:00 92.8 F 0.1  
6:00 109.1 F 1.2  
8:00 142.7 F 3.7  
10:00 167.1 F 7.5  
12:00 186.6 F 11.5  
14:00 204.9 F 2.9  
16:00 212.0 F 6.6  
18:00 223.2 F 10.2  
20:00 231.5 F 8.9  
22:00 237.8 F 10.8  
24:00 234.9 F 8.3  
26:00 248.6 F 14.9  
28:00 260.0 F 21.3  
30:00 270.7 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.4 F 29.4  
2:00 271.5 F 29.0  
3:00 272.4 F 29.2  
4:00 271.5 F 29.1

Min 271.4 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
5:00 230.9 F 0.0  
10:00 193.1 F 0.0  
15:00 183.4 F 0.0  
20:00 180.0 F 0.0  
25:00 177.3 F 0.0  
30:00 175.8 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:48  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:49  
CYCLE COMPLETE

02/15/2024 11:01 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_