
Midmark M11 - v1.0.5
Total Cycles: 3536

Name: RITTER2
Sterilizer ID: V2313228
04/17/2024 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.9 F 0.0
2:00 159.0 F 2.4
4:00 165.2 F 5.9
6:00 191.0 F 9.5
8:00 208.9 F 6.7
10:00 220.2 F 8.4
12:00 233.5 F 13.2
14:00 235.4 F 10.4
16:00 235.4 F 8.5
18:00 249.0 F 15.5
20:00 260.9 F 22.2
22:00 271.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.1
2:00 272.1 F 29.5
3:00 271.1 F 29.3
4:00 271.8 F 30.1

Min 271.1 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 205.0 F 0.0
10:00 214.5 F 0.0
15:00 210.5 F 0.0
20:00 209.6 F 0.0
25:00 211.0 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

04/17/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____