
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000492
9:33 16/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:23
STERILIZING 4:49
136.0 C 228kPa 4:49
136.4 C 233kPa 5:19
136.5 C 230kPa 5:49
136.3 C 232kPa 6:19
136.4 C 233kPa 6:49
136.4 C 233kPa 7:19
136.0 C 230kPa 7:49
136.4 C 232kPa 8:19
136.4 C 232kPa 8:19
VENTING 8:19
AIR DRYING 9:13
CYCLE COMPLETE 24:13

Digital Signature #
6AC8DD713E8A4AAC
