

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1778

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2022 01:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.7 F 0.0  
2:00 168.6 F 2.9  
4:00 181.2 F 6.4  
6:00 209.6 F 11.0  
8:00 217.4 F 5.6  
10:00 230.6 F 11.0  
12:00 231.1 F 7.7  
14:00 242.0 F 11.8  
16:00 243.8 F 12.7  
18:00 256.6 F 19.5  
20:00 268.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.6 F 29.6  
2:00 271.7 F 28.8  
3:00 272.2 F 29.6  
4:00 272.4 F 29.3

Min 271.4 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 215.9 F 0.0  
10:00 227.1 F 0.0  
15:00 236.4 F 0.0  
20:00 229.8 F 0.0  
25:00 227.5 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:13  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

03/31/2022 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_