
Midmark M11 - v1.0.5
Total Cycles: 1641

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 77.7 F 0.0
4:00 81.6 F 0.0
6:00 104.8 F 0.1
8:00 141.3 F 0.2
10:00 173.3 F 2.1
12:00 196.1 F 5.4
14:00 208.9 F 9.1
16:00 224.3 F 12.9
18:00 227.8 F 7.5
20:00 238.6 F 11.9
22:00 235.6 F 9.1
24:00 241.1 F 10.6
26:00 245.9 F 13.3
28:00 257.2 F 19.4
30:00 267.6 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.4 F 29.3
2:00 271.4 F 28.5
3:00 272.3 F 29.4
4:00 272.0 F 28.8

Min 271.1 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 225.7 F 0.0
10:00 226.1 F 0.0
15:00 222.6 F 0.0
20:00 227.0 F 0.0
25:00 230.8 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

10/24/2022 08:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____