
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000734
7:39 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:22
STERILIZING 3:16
136.0 C 227kPa 3:16
136.4 C 232kPa 3:46
136.4 C 233kPa 4:16
136.3 C 231kPa 4:46
136.4 C 233kPa 5:16
136.4 C 233kPa 5:46
136.4 C 232kPa 6:16
136.4 C 232kPa 6:46
136.4 C 232kPa 6:46
VENTING 6:46
AIR DRYING 7:34
CYCLE COMPLETE 7:43

Digital Signature #
CF9AC9ACBD8C9782
