
Midmark M11 - v1.0.0
Total Cycles: 1557

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2022 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 98.1 F 0.0
4:00 91.8 F 0.3
6:00 119.5 F 2.4
8:00 148.7 F 5.0
10:00 170.3 F 8.0
12:00 198.9 F 11.7
14:00 208.7 F 3.1
16:00 219.6 F 6.9
18:00 230.9 F 10.6
20:00 230.9 F 7.5
22:00 240.5 F 11.7
24:00 238.0 F 9.5
26:00 251.1 F 15.9
28:00 262.3 F 22.4
30:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.8 F 29.2
2:00 271.5 F 28.3
3:00 272.4 F 29.3
4:00 272.2 F 28.8

Min 271.2 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 230.9 F 0.0
10:00 209.6 F 0.0
15:00 217.1 F 0.0
20:00 222.9 F 0.0
25:00 219.1 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:38
CYCLE COMPLETE

01/25/2022 09:05 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____