
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002308
11:09 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:08
136.4 C 222kPa 3:08
136.4 C 226kPa 3:38
136.3 C 224kPa 4:08
136.4 C 225kPa 4:38
136.4 C 224kPa 5:08
136.3 C 225kPa 5:38
136.4 C 222kPa 6:08
136.3 C 225kPa 6:38
136.3 C 225kPa 6:38
VENTING 6:38
AIR DRYING 7:27
CYCLE COMPLETE 9:17

Digital Signature #
7CA6EE32A8265142
