
Midmark M11 - v1.0.0
Total Cycles: 1711

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.2 F 0.0
2:00 179.8 F 2.7
4:00 191.7 F 6.8
6:00 211.9 F 12.1
8:00 223.2 F 6.6
10:00 234.8 F 11.8
12:00 235.3 F 8.9
14:00 240.1 F 9.8
16:00 247.7 F 13.9
18:00 260.2 F 20.9
20:00 271.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.9
2:00 271.4 F 28.1
3:00 272.2 F 28.9
4:00 272.1 F 28.5

Min 271.2 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 221.3 F 0.0
10:00 233.5 F 0.0
15:00 234.5 F 0.0
20:00 219.0 F 0.0
25:00 219.3 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

03/14/2022 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____