

-----  
Midmark M11 - v1.0.0  
Total Cycles: 663

Name: MIDMARK  
Sterilizer ID: V2178198  
02/10/2021 02:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
2:00 190.8 F 4.7  
4:00 217.3 F 11.8  
6:00 229.7 F 8.9  
8:00 233.0 F 8.3  
10:00 240.7 F 10.6  
12:00 249.0 F 14.6  
14:00 263.3 F 22.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 271.7 F 28.1  
2:00 272.2 F 28.3  
3:00 272.8 F 28.7  
4:00 271.9 F 28.6  
  
Min 271.2 F 28.0  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 232.4 F 0.0  
10:00 237.3 F 0.0  
15:00 205.9 F 0.0  
20:00 192.9 F 0.0  
25:00 185.2 F 0.0  
30:00 181.0 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 15:48  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 00:55:07  
CYCLE COMPLETE

02/10/2021 03:24 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01