
Midmark M11 - v1.0.5
Total Cycles: 4018

Name: RITTER2
Sterilizer ID: V2313228
10/10/2024 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 159.3 F 1.9
4:00 161.4 F 5.2
6:00 181.9 F 9.2
8:00 202.1 F 13.0
10:00 219.1 F 7.5
12:00 231.3 F 12.1
14:00 231.8 F 9.0
16:00 239.3 F 10.6
18:00 244.8 F 13.5
20:00 256.5 F 19.8
22:00 267.3 F 26.5
24:00 273.3 F 30.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 272.8 F 30.5
2:00 271.7 F 30.1
3:00 272.7 F 30.5
4:00 271.7 F 29.9

Min 271.6 F 29.9
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 212.7 F 0.0
10:00 222.0 F 0.0
15:00 222.7 F 0.0
20:00 223.7 F 0.0
25:00 228.4 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:20
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

10/10/2024 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____