

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5127

Name: RITTER 1  
Sterilizer ID: V2178198  
11/22/2024 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.9 F 0.0  
2:00 174.2 F 0.0  
4:00 184.7 F 0.4  
6:00 205.8 F 1.2  
8:00 218.1 F 5.0  
10:00 226.9 F 8.8  
12:00 237.2 F 13.2  
14:00 236.7 F 10.0  
16:00 238.4 F 9.2  
18:00 245.2 F 12.5  
20:00 245.2 F 12.9  
22:00 257.0 F 19.1  
24:00 267.6 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.2 F 28.9  
2:00 272.5 F 28.9  
3:00 272.0 F 28.5  
4:00 271.5 F 28.2

Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 202.6 F 0.0  
10:00 209.0 F 0.0  
15:00 224.9 F 0.0  
20:00 216.6 F 0.0  
25:00 215.8 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 25:33  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:04:14  
CYCLE COMPLETE

11/22/2024 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_