
Midmark M11 - v1.0.5
Total Cycles: 1337

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.8 F 0.0
2:00 185.4 F 1.8
4:00 197.6 F 6.9
6:00 225.3 F 12.4
8:00 229.5 F 8.9
10:00 236.2 F 10.0
12:00 242.9 F 13.2
14:00 243.0 F 12.5
16:00 256.0 F 19.4
18:00 267.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 29.8
2:00 271.3 F 29.2
3:00 272.3 F 30.0
4:00 271.8 F 29.4

Min 271.2 F 29.1
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 229.8 F 0.0
10:00 244.6 F 0.0
15:00 243.3 F 0.0
20:00 227.9 F 0.0
25:00 226.8 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

07/25/2022 12:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____