
Midmark M11 - v1.0.0
Total Cycles: 2518

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 192.0 F 3.2
4:00 206.7 F 6.9
6:00 208.2 F 12.5
8:00 223.0 F 6.6
10:00 237.0 F 11.7
12:00 234.5 F 8.6
14:00 241.3 F 10.3
16:00 246.5 F 13.6
18:00 258.6 F 20.3
20:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.9
2:00 272.2 F 29.2
3:00 271.9 F 28.6
4:00 272.5 F 29.3

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 234.8 F 0.0
10:00 231.9 F 0.0
15:00 228.8 F 0.0
20:00 226.5 F 0.0
25:00 230.4 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

11/07/2022 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____