
Midmark M11 - v1.0.0
Total Cycles: 1862

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 170.8 F 3.2
4:00 183.4 F 7.3
6:00 201.4 F 12.4
8:00 220.2 F 6.7
10:00 232.2 F 11.9
12:00 234.1 F 8.6
14:00 241.2 F 10.4
16:00 246.5 F 13.5
18:00 259.0 F 20.3
20:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.4 F 29.2
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 271.7 F 28.3

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 226.8 F 0.0
10:00 226.9 F 0.0
15:00 241.3 F 0.0
20:00 233.6 F 0.0
25:00 229.7 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

04/25/2022 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____