
Midmark M11 - v1.0.5
Total Cycles: 2941

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.6 F 0.0
4:00 77.7 F 0.0
6:00 109.1 F 0.0
8:00 147.6 F 0.6
10:00 169.5 F 2.3
12:00 195.2 F 5.5
14:00 225.9 F 9.5
16:00 228.4 F 13.5
18:00 230.2 F 7.9
20:00 241.3 F 12.6
22:00 237.4 F 9.7
24:00 238.0 F 9.3
26:00 248.5 F 14.5
28:00 260.0 F 20.9
30:00 270.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 28.9
2:00 271.9 F 29.0
3:00 272.1 F 28.6
4:00 272.4 F 29.3

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 237.5 F 0.0
10:00 224.0 F 0.0
15:00 219.9 F 0.0
20:00 229.0 F 0.0
25:00 228.7 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

09/28/2023 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____