

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4214

Name: RITTER 1  
Sterilizer ID: V2178198  
01/19/2024 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.9 F 0.0  
2:00 170.0 F 1.5  
4:00 172.2 F 4.6  
6:00 206.1 F 8.8  
8:00 212.4 F 13.5  
10:00 221.2 F 6.8  
12:00 232.0 F 11.3  
14:00 233.4 F 7.8  
16:00 244.9 F 13.5  
18:00 243.9 F 12.0  
20:00 255.4 F 18.1  
22:00 266.1 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.4 F 28.5  
2:00 272.0 F 28.9  
3:00 271.8 F 28.6  
4:00 272.3 F 28.6

Min 271.0 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 227.0 F 0.0  
10:00 242.8 F 0.0  
15:00 224.1 F 0.0  
20:00 225.5 F 0.0  
25:00 194.9 F 0.0  
30:00 187.2 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 23:44  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:02:56  
CYCLE COMPLETE

01/19/2024 11:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_