
Midmark M11 - v1.0.5
Total Cycles: 338

Name: RITTER2
Sterilizer ID: V2313228
09/08/2021 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 185.3 F 2.7
4:00 197.7 F 7.6
6:00 216.0 F 12.6
8:00 225.0 F 8.6
10:00 235.9 F 11.3
12:00 241.0 F 12.5
14:00 241.6 F 11.8
16:00 254.4 F 18.5
18:00 265.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.5 F 29.6
2:00 272.9 F 30.2
3:00 272.0 F 29.2
4:00 272.9 F 29.9

Min 271.8 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 201.6 F 0.0
10:00 212.5 F 0.0
15:00 212.0 F 0.0
20:00 209.7 F 0.0
25:00 196.0 F 0.0
30:00 197.0 F 0.1
Drying Complete

FILLING: 2:05
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

09/08/2021 10:22 AM
ACCEPTED: X
REJECTED:
Operator: U01