
Midmark M11 - v1.0.0
Total Cycles: 3994

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2023 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 179.8 F 2.3
4:00 183.9 F 6.8
6:00 205.1 F 11.8
8:00 219.6 F 5.6
10:00 232.2 F 10.6
12:00 232.2 F 7.2
14:00 245.5 F 13.1
16:00 243.1 F 11.5
18:00 255.8 F 18.1
20:00 267.1 F 25.0
22:00 272.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.9 F 28.8
2:00 271.8 F 28.6
3:00 272.8 F 28.7
4:00 271.6 F 28.5

Min 271.4 F 28.2
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 217.6 F 0.0
10:00 205.7 F 0.0
15:00 198.3 F 0.0
20:00 197.9 F 0.0
25:00 198.1 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

11/08/2023 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____