

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4015

Name: RITTER 1  
Sterilizer ID: V2178198  
11/13/2023 05:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.9 F 0.0  
2:00 135.6 F 2.1  
4:00 152.8 F 5.4  
6:00 192.1 F 9.0  
8:00 213.7 F 8.1  
10:00 214.8 F 6.6  
12:00 225.4 F 10.9  
14:00 229.0 F 6.6  
16:00 240.7 F 12.1  
18:00 239.0 F 10.1  
20:00 251.8 F 16.2  
22:00 262.7 F 22.6  
24:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.0  
2:00 271.3 F 28.3  
3:00 272.8 F 29.1  
4:00 271.8 F 28.5

Min 271.3 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
5:00 226.6 F 0.0  
10:00 230.5 F 0.0  
15:00 224.5 F 0.0  
20:00 224.5 F 0.0  
25:00 224.8 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:03:11  
CYCLE COMPLETE

11/13/2023 06:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_