
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002399
9:17 20/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:20
STERILIZING 3:23
136.4 C 225kPa 3:23
136.2 C 223kPa 3:53
136.4 C 225kPa 4:23
136.4 C 225kPa 4:53
136.4 C 225kPa 5:23
136.4 C 225kPa 5:53
136.4 C 225kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:44
CYCLE COMPLETE 37:44

Digital Signature #
F9590EDA0F1F4621
