
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002823
9:13 13/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:25
STERILIZING 4:32
136.4 C 223kPa 4:32
136.2 C 222kPa 5:02
136.4 C 222kPa 5:32
136.4 C 224kPa 6:02
136.1 C 222kPa 6:32
136.2 C 222kPa 7:02
136.1 C 219kPa 7:32
135.9 C 218kPa 8:02
135.9 C 218kPa 8:02
VENTING 8:02
AIR DRYING 8:52
CYCLE COMPLETE 8:53

Digital Signature #
D9FF7290C1079EA7
