
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004447
9:11 27/09/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:39
PRESSURIZING 1:49
STERILIZING 2:47
136.0 C 220kPa 2:47
136.3 C 223kPa 3:17
136.4 C 223kPa 3:47
136.4 C 223kPa 4:17
136.0 C 221kPa 4:47
136.1 C 220kPa 5:17
136.4 C 223kPa 5:47
136.3 C 223kPa 6:17
136.3 C 223kPa 6:17
VENTING 6:17
AIR DRYING 7:08
CYCLE COMPLETE 7:14

Digital Signature #
3F1ED11EA94A2336
