
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000035
15:52 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 3:56
STERILIZING 4:55
136.0 C 222kPa 4:55
136.4 C 224kPa 5:25
136.0 C 224kPa 5:55
136.3 C 225kPa 6:25
136.5 C 228kPa 6:55
136.4 C 225kPa 7:25
136.4 C 227kPa 7:55
136.4 C 226kPa 8:25
VENTING 8:25
AIR DRYING 9:16
CYCLE COMPLETE 11:17

Digital Signature #
078F53340F9BFCB1
