

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2825

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2023 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.3 F 0.0  
2:00 130.1 F 0.0  
4:00 160.3 F 0.0  
6:00 198.3 F 1.3  
8:00 208.0 F 5.4  
10:00 224.3 F 9.7  
12:00 234.5 F 9.1  
14:00 238.8 F 11.0  
16:00 236.0 F 9.0  
18:00 239.8 F 10.0  
20:00 248.3 F 14.5  
22:00 260.4 F 21.3  
24:00 271.5 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.5 F 29.0  
2:00 271.7 F 28.9  
3:00 272.3 F 28.9  
4:00 271.6 F 28.9

Min 271.4 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.7 F 0.0  
5:00 239.2 F 0.0  
10:00 219.5 F 0.0  
15:00 205.7 F 0.0  
20:00 202.7 F 0.0  
25:00 202.1 F 0.0  
30:00 201.6 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:02:49  
CYCLE COMPLETE

08/31/2023 03:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_