

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5285

Name: RITTER 1  
Sterilizer ID: V2178198  
01/28/2025 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 185.0 F 2.5  
4:00 188.1 F 6.2  
6:00 198.5 F 10.1  
8:00 226.1 F 7.1  
10:00 221.2 F 7.4  
12:00 233.5 F 11.8  
14:00 233.1 F 8.4  
16:00 244.8 F 12.3  
18:00 243.1 F 12.2  
20:00 254.8 F 18.1  
22:00 265.0 F 24.3  
24:00 271.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 271.9 F 29.3  
2:00 272.5 F 29.6  
3:00 272.4 F 29.3  
4:00 272.3 F 29.1

Min 271.3 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 210.6 F 0.0  
10:00 212.0 F 0.0  
15:00 211.7 F 0.0  
20:00 214.3 F 0.0  
25:00 216.0 F 0.0  
30:00 213.9 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 24:08  
STERILIZING: 4:00  
VENTING: 4:00  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

01/28/2025 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_