
Midmark M11 - v1.0.5
Total Cycles: 4242

Name: RITTER2
Sterilizer ID: V2313228
01/06/2025 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.9 F 0.0
2:00 72.1 F 0.0
4:00 123.5 F 0.0
6:00 110.9 F 0.1
8:00 147.5 F 2.1
10:00 166.1 F 4.9
12:00 183.4 F 8.2
14:00 200.0 F 11.8
16:00 210.9 F 4.3
18:00 223.0 F 8.0
20:00 233.3 F 12.2
22:00 233.5 F 8.9
24:00 241.9 F 11.4
26:00 244.8 F 13.0
28:00 256.3 F 19.1
30:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.1
2:00 272.7 F 29.6
3:00 271.8 F 28.9
4:00 272.5 F 29.2

Min 271.6 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 218.4 F 0.0
10:00 219.9 F 0.0
15:00 211.3 F 0.0
20:00 201.8 F 0.0
25:00 210.2 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

01/06/2025 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____