
Midmark M11 - v1.0.5
Total Cycles: 1771

Name: RITTER2
Sterilizer ID: V2313228
11/23/2022 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 220.4 F 0.0
2:00 215.0 F 4.4
4:00 211.3 F 10.8
6:00 226.3 F 7.5
8:00 238.1 F 11.8
10:00 243.0 F 13.0
12:00 244.4 F 12.8
14:00 258.4 F 20.4
16:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.6 F 29.6
2:00 271.9 F 29.0
3:00 271.8 F 29.4
4:00 272.5 F 29.6

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 227.3 F 0.0
10:00 238.3 F 0.0
15:00 228.5 F 0.0
20:00 225.6 F 0.0
25:00 225.1 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 2:44
HEATING: 16:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:56:39
CYCLE COMPLETE

11/23/2022 03:54 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____