
Midmark M11 - v1.0.5
Total Cycles: 2525

Name: RITTER2
Sterilizer ID: V2313228
06/16/2023 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 76.6 F 0.0
4:00 98.9 F 0.0
6:00 138.0 F 0.0
8:00 172.8 F 0.3
10:00 195.4 F 3.5
12:00 221.9 F 7.4
14:00 229.8 F 11.4
16:00 227.4 F 5.9
18:00 233.9 F 10.3
20:00 236.8 F 8.7
22:00 244.3 F 13.1
24:00 242.9 F 11.9
26:00 255.5 F 18.5
28:00 267.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.8 F 29.0
2:00 272.3 F 29.3
3:00 272.7 F 29.5
4:00 272.8 F 29.5

Min 271.7 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 232.8 F 0.0
Dry Cycle Aborted

FILLING: 1:02
HEATING: 29:16
STERILIZING: 4:00
VENTING: 2:44
DRYING: 1:51
TOTAL CYCLE: 00:38:54

06/16/2023 09:08 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____