
Midmark M11 - v1.0.5
Total Cycles: 2849

Name: RITTER2
Sterilizer ID: V2313228
09/08/2023 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 71.6 F 0.0
4:00 76.1 F 0.0
6:00 102.2 F 0.0
8:00 139.1 F 0.4
10:00 171.3 F 3.2
12:00 190.4 F 6.7
14:00 201.8 F 10.7
16:00 216.0 F 6.6
18:00 223.1 F 8.7
20:00 235.3 F 13.1
22:00 235.6 F 9.9
24:00 236.6 F 9.2
26:00 246.5 F 14.0
28:00 257.9 F 20.2
30:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 29.1
3:00 272.5 F 29.8
4:00 271.5 F 29.0

Min 271.4 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 195.8 F 0.0
10:00 208.2 F 0.0
15:00 200.0 F 0.0
20:00 204.5 F 0.0
25:00 208.0 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

09/08/2023 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____