

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002377  
12:00 13/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:22  
STERILIZING 3:23  
136.4 C 220kPa 3:23  
136.4 C 224kPa 3:53  
136.4 C 223kPa 4:23  
136.4 C 226kPa 4:53  
136.1 C 224kPa 5:23  
136.3 C 223kPa 5:53  
136.3 C 225kPa 6:23  
136.4 C 226kPa 6:53  
136.4 C 226kPa 6:53  
VENTING 6:53  
AIR DRYING 7:49  
CYCLE COMPLETE 37:49

-----  
Digital Signature #  
37951DB0042F9D42  
-----