
Midmark M11 - v1.0.0
Total Cycles: 1196

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2021 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.0 F 0.0
2:00 120.4 F 0.0
4:00 150.7 F 0.3
6:00 177.6 F 3.5
8:00 196.1 F 7.2
10:00 211.3 F 11.3
12:00 219.0 F 4.8
14:00 232.0 F 9.7
16:00 237.1 F 9.8
18:00 242.5 F 12.1
20:00 240.7 F 10.6
22:00 253.6 F 17.1
24:00 264.8 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.9
2:00 272.1 F 28.4
3:00 271.4 F 28.1
4:00 272.6 F 29.0

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 202.4 F 0.0
10:00 194.0 F 0.0
15:00 192.3 F 0.0
20:00 188.7 F 0.0
25:00 209.8 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:58
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

09/22/2021 02:17 PM
ACCEPTED: X
REJECTED:
Operator: U01