

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2566

Name: RITTER2  
Sterilizer ID: V2313228  
06/26/2023 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
2:00 194.7 F 3.9  
4:00 208.0 F 9.5  
6:00 222.9 F 7.4  
8:00 235.1 F 11.5  
10:00 235.6 F 9.6  
12:00 237.1 F 9.1  
14:00 249.7 F 15.7  
16:00 262.2 F 22.8  
18:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.6  
2:00 271.4 F 28.9  
3:00 272.8 F 29.8  
4:00 272.0 F 29.2  
  
Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 220.6 F 0.0  
10:00 232.8 F 0.0  
15:00 229.3 F 0.0  
20:00 228.0 F 0.0  
25:00 229.4 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

06/26/2023 12:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_