
Midmark M11 - v1.0.0
Total Cycles: 2490

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2022 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 172.8 F 2.4
4:00 195.1 F 6.5
6:00 211.8 F 11.2
8:00 219.1 F 3.9
10:00 229.2 F 8.8
12:00 240.3 F 12.9
14:00 238.1 F 10.5
16:00 237.0 F 7.9
18:00 248.7 F 14.8
20:00 260.3 F 21.2
22:00 270.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.7 F 28.5
3:00 271.8 F 29.1
4:00 272.2 F 29.1

Min 271.2 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 223.1 F 0.0
10:00 235.8 F 0.0
15:00 235.8 F 0.0
20:00 229.9 F 0.0
25:00 231.0 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

10/26/2022 10:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____