
Midmark M11 - v1.0.0
Total Cycles: 731

Name: RITTER 1
Sterilizer ID: V2178198
03/18/2021 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.1 F 0.0
2:00 162.0 F 3.4
4:00 186.6 F 8.1
6:00 209.0 F 13.4
8:00 220.4 F 8.9
10:00 234.9 F 8.8
12:00 243.5 F 12.8
14:00 241.6 F 11.3
16:00 255.1 F 18.5
18:00 267.8 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 271.7 F 28.6
2:00 272.2 F 28.9
3:00 272.2 F 29.7
4:00 272.0 F 29.2

Min 271.0 F 28.6
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 221.5 F 0.0
10:00 226.8 F 0.0
15:00 227.0 F 0.0
20:00 231.4 F 0.0
25:00 233.6 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:11
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

03/18/2021 12:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____