

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000941  
17:31 17/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:46  
STERILIZING 3:55  
136.0 C 228kPa 3:55  
136.0 C 229kPa 4:25  
136.0 C 229kPa 4:55  
136.3 C 233kPa 5:25  
135.9 C 230kPa 5:55  
136.4 C 233kPa 6:25  
136.1 C 229kPa 6:55  
136.4 C 232kPa 7:25  
136.4 C 232kPa 7:25  
VENTING 7:25  
AIR DRYING 8:18  
CYCLE COMPLETE 23:18  
-----

Digital Signature #  
37CFD21FA75FEE3C  
-----