
Midmark M11 - v1.0.0
Total Cycles: 4456

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 167.0 F 1.9
4:00 167.0 F 5.2
6:00 185.6 F 9.1
8:00 202.8 F 13.4
10:00 214.9 F 5.6
12:00 228.8 F 9.8
14:00 234.8 F 9.5
16:00 240.4 F 11.5
18:00 237.8 F 9.6
20:00 250.7 F 15.7
22:00 261.5 F 21.8
24:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.3
2:00 272.0 F 28.7
3:00 271.2 F 28.6
4:00 272.6 F 29.3

Min 271.1 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 216.6 F 0.0
10:00 220.1 F 0.0
15:00 219.1 F 0.0
20:00 224.8 F 0.0
25:00 230.5 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

04/15/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____