

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3765

Name: RITTER 1  
Sterilizer ID: V2178198  
09/13/2023 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.6 F 0.0  
2:00 148.8 F 0.0  
4:00 163.6 F 1.7  
6:00 195.5 F 5.2

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 207F  
CHAMBER PRESS 9.5PSI  
\*\* ITEMS NOT STERILE \*\*

09/13/2023 10:48 AM